

BIOHIT COLONVIEW

quick test




Ordering details:

Cat. No.	Item
602 250.02	30 tests/box

Product specifications

- Sample: stool sample
- Test principle: immunological detection of Hb and Hb/Hp complex
- Kit components: sampling tube, sampling paper, test cassette, instructions for use
- Storage: +2-30°C



In the USA and Japan for research use only.

Biohit Oyj

Laippatie 1
00880 Helsinki
Finland
Tel. +358 9 773 861
Fax +358 9 773 86 200
info@biohit.fi

Biohit HealthCare Srl

Via Boncompagni, 3
20139 Milano
Italy
Tel +39 02 38238113
Fax +39 02 38236521
info.italy@biohit.fi

Biohit HealthCare Ltd.

Pioneer House, North Rd
Ellesmere Port, CH65 1AD,
United Kingdom
Tel. +44 151 550 4 550
Fax. +44 151 550 4 551
info@biohithealthcare.co.uk

www.biohithealthcare.com

BIOHIT HealthCare

Innovating for Health



VERSATILE SCREENING TOOL FOR COLORECTAL CANCER

490089EN 06/2014



BIOHIT COLONVIEW

quick test

Detects bleeding derived from both lower and upper GI tract

Detection of occult blood in feces is essential in the early diagnosis of colorectal cancer. The Biohit ColonView quick test is an immunological lateral flow test which detects hemoglobin and hemoglobin-haptoglobin complex (degraded hemoglobin) from stool. With this combination of two biomarkers, the ColonView test can be used for detecting bleeding derived from the colon or upper parts of the GI tract. This makes ColonView a superior tool to detect bleeding cancer tumors and their precursors.

- Detects bleeding from lower and upper GI tract
- Superior sensitivity and PPV
- Hygienic sampling, transportation and analysis
- No dietary restrictions before sample collection
- Results in 15 min

One sample for dual testing

Clever design of the ColonView sampling tube eliminates patient-errors and ensures that there will always be a correct sample amount. The sampling tube contains all the necessary reagents and no additional accessories are needed for the analysis. One sample is enough for the testing of both Hb and Hb/Hp complex, and the result is ready in 15 minutes.

Testing made easy

Test performance

The stool specimens are collected easily with the stool collection paper and sample applicator stick:

1. Make bowel movement onto the collection paper


2. Unscrew the blue end and remove the sample applicator

3. Collect sample by inserting the stick into the stool specimen

4. Re-insert applicator stick

5. Unscrew the white cap and drop three drops to the test cassette

6. Read after 15 minutes



Improved accuracy to the screening of colorectal cancer

